

A Multi-Site Study Comparing an 18-24h Sensititre® Susceptibility System to the CLSI Broth Microdilution Method for Telavancin Using Fastidious and Non-Fastidious Gram-Positive Organisms

*N. M. Holliday¹, C. M. Bastulli¹, C. C. Knapp¹, S. B. Killian¹, K. M. Krause², J. M. Streit³, R. N. Jones³, P. R. Rhomberg³, L. Define⁴, J. Beach⁴, J. DiPersio⁴
¹TREK Diagnostic Systems, Cleveland, OH; ²Theravance Inc., South San Francisco, CA; ³JMI Lab., North Liberty, IA; ⁴Summa Health Systems, Akron, OH

Nicole Holliday
TREK Diagnostic Systems
Lab Services
982 Keynote Circle, Suite 6
Cleveland, Ohio 44131
Phone: 1-800-571-8899
Email: nholliday@trekds.com

Abstract

Background: Telavancin (TLV) (Theravance, Inc., South San Francisco, CA) is a bactericidal lipopeptide under investigation for the treatment of cSSSI and HAP caused by methicillin-resistant *Staphylococcus aureus* (MRSA) and other Gram-positive bacteria. An evaluation was performed to determine the accuracy and reproducibility of TLV susceptibility testing using the Sensititre® 18-24h dried susceptibility system (TREK Diagnostic Systems, Cleveland, OH) compared with the CLSI M07 reference broth microdilution method (BMD). Both automated and manual reading methods were performed.

Methods: TLV (0.001-16 µg/mL) was tested against 897 recent clinical isolates, 150 challenge isolates and 50 reproducibility isolates. These isolates consisted of: 137 CNS spp., 186 *Streptococcus aureus*, 128 *Enterococcus* spp., 326 *Streptococcus* spp., Beta-hemolytic gr., 186 *Streptococcus pneumoniae*, and 134 *Streptococcus* spp. Viridans gr. Dried plates were inoculated as per manufacturers' instructions and BMD was performed per CLSI M07. Recommended CLSI quality control (QC) organisms were tested daily and all results were within the CLSI published QC ranges.

Results: Comparisons of TLV MIC results on the Sensititre® system to the CLSI M07 BMD for both automated and manual results resulted in 99.0% and 99.6% essential agreement (+/- one log₂ dilution), respectively. Overall agreement for the reproducibility (+/- one log₂ dilution of the modal MIC) for automated and manual reads were 99.5% and 98.5%, respectively.

Conclusions: The results for TLV indicate that the Sensititre® 18-24h susceptibility system for all clinical and challenge isolates gives reliable results using either the automated or manual read methods compared to the reference CLSI BMD.

Introduction

Telavancin is a bactericidal lipopeptide antibiotic structure derived from vancomycin. It is highly active against Gram-positive bacteria including methicillin-resistant *Staphylococcus aureus* (MRSA). Here we are reporting results from a multi-site study, performing a series of evaluations to determine the accuracy and reproducibility of the Sensititre® 18-24h susceptibility system with telavancin compared to the CLSI reference broth microdilution method (M07).

Materials & Methods

• Indications for use: The Sensititre® 18-24 hour MIC or breakpoint susceptibility system is an *in vitro* diagnostic product for clinical susceptibility testing of both fastidious and non-fastidious organisms.

• Each isolate was tested using a Sensititre® 18-24 hour susceptibility plate containing approximately 447 Gram-positive isolates and 450 fastidious isolates from each site. Also included were 75 Gram-positive and 75 fastidious challenge isolates from the Center for Disease Control and Prevention (CDC) that were supplied to a single testing site (Tables 1 and 2).

• The CLSI reference broth microdilution plate was prepared and tested on each isolate according to the Clinical Laboratory Standards Institute (CLSI M07).

• Testing consisted of 897 fresh clinical isolates (conducted at three different sites) including approximately 447 Gram-positive isolates and 450 fastidious isolates from each site. Also included were 75 Gram-positive and 75 fastidious challenge isolates from the Center for Disease Control and Prevention (CDC) that were supplied to a single testing site (Tables 1 and 2).

• Reproducibility testing consisted of 25 Gram-positive and 25 *Streptococcus* spp. isolates tested at all 3 sites on the Sensititre® 18-24 hour susceptibility plate (Table 1). The test plate results were compared with those of the CLSI reference broth microdilution plate.

• Quality control (QC) was assured by testing 20 replicates of each ATCC strain including *S. aureus* 29213, *E. faecalis* 29212, and *S. pneumoniae* 49619, at each of the three sites (Tables 1 and 3).

• Colony counts were performed on the inoculum of the QC strains on each day of testing.

Organisms Tested	Number Tested
Clinical isolates (3 sites) (447 Gram-positive, 450 <i>Streptococcus</i> spp.)	897
CDC Challenge Isolates (one site) (75 Gram-positive, 75 <i>Streptococcus</i> spp.)	150
Reproducibility Isolates (3 sites) (25 Gram-positive, 25 <i>Streptococcus</i> spp.)	50
CLSI Quality Control Strains (20 replicates of each strain at 3 sites)	3 x 20

Gram-positive Organisms	Number Tested
<i>Coagulase Negative Staphylococcus</i>	177
<i>Staphylococcus aureus</i>	121
<i>Enterococcus</i> spp.	130
Beta <i>Streptococcus</i> (MHB)	90
Total	518
<i>Streptococcus</i> spp.	Number Tested
<i>Streptococcus pneumoniae</i>	176
Viridans Group <i>Streptococcus</i>	130
Beta <i>Streptococcus</i> (MHB with LHB)	220
Total	526

Quality Control Strains	CLSI MIC Ranges (µg/ml)
<i>Staphylococcus aureus</i> ATCC 29213	0.12-1
<i>Enterococcus faecalis</i> ATCC 29212	0.12-0.5
<i>Streptococcus pneumoniae</i> ATCC 49619	0.004-0.03

Table 4. Summary Data and % Essential Agreement of Gram-positive Clinical and Challenge Isolates Using the Manual Read Method

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Coagulase Negative Staphylococcus</i>	113	112	113	112	100.0%	100.0%
<i>Staphylococcus aureus</i>	150	150	150	150	100.0%	100.0%
<i>Enterococcus</i> spp.	105	101	103	99	98.1%	98.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with MHB	79	50	77	48	97.5%	98.0%
Total	447	413	443	409	99.1%	99.0%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Coagulase Negative Staphylococcus</i>	17	17	17	17	100.0%	100.0%
<i>Staphylococcus aureus</i>	27	27	27	27	100.0%	100.0%
<i>Enterococcus</i> spp.	16	15	16	15	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with MHB	11	6	11	6	100%	100%
Total	71	65	71	65	100%	100%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Coagulase Negative Staphylococcus</i>	130	129	130	129	100.0%	100.0%
<i>Staphylococcus aureus</i>	177	177	177	177	100.0%	100.0%
<i>Enterococcus</i> spp.	121	116	121	116	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with MHB	90	58	88	54	97.8%	98.4%
Total	518	478	516	476	99.6%	99.6%

Table 5. Summary Data and % Essential Agreement of Gram-positive Clinical and Challenge Isolates Using the Auto Read Method

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Coagulase Negative Staphylococcus</i>	112	111	111	110	99.1%	99.1%
<i>Staphylococcus aureus</i>	150	150	150	150	100%	100%
<i>Enterococcus</i> spp.	104	101	102	99	98.1%	98.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with MHB	78	44	77	43	98.7%	97.7%
Total	444	406	440	402	99.1%	99.0%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Coagulase Negative Staphylococcus</i>	17	17	17	17	100.0%	100.0%
<i>Staphylococcus aureus</i>	27	27	27	27	100.0%	100.0%
<i>Enterococcus</i> spp.	16	15	15	14	93.8%	93.3%
<i>Streptococcus</i> spp. Beta-hemolytic Group with MHB	11	8	11	8	100%	100%
Total	71	67	70	66	98.6%	98.5%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Coagulase Negative Staphylococcus</i>	129	128	128	127	99.0%	99.0%
<i>Staphylococcus aureus</i> ATCC 29213	177	177	177	177	100%	100%
<i>Enterococcus</i> spp.	120	116	117	113	96.0%	96.7%
<i>Streptococcus</i> spp. Beta-hemolytic Group with MHB	89	52	88	51	99.4%	98.3%
Total	515	473	510	468	99.0%	98.9%

Table 6. Summary Data and % Essential Agreement of *Streptococcus* spp. Clinical and Challenge Isolates Using the Manual Read Method

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Streptococcus pneumoniae</i>	151	151	150	150	99.3%	99.3%
<i>Streptococcus</i> spp. Viridans Group	104	104	104	104	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with LHB	195	194	194	193	99.5%	99.5%
Total	450	449	448	447	99.8%	99.6%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Streptococcus pneumoniae</i>	25	25	25	25	100.0%	100.0%
<i>Streptococcus</i> spp. Viridans Group	26	25	26	25	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with LHB	25	25	25	25	100.0%	100.0%
Total	76	75	76	75	100.0%	100.0%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Streptococcus pneumoniae</i>	178	178	175	175	99.7%	99.7%
<i>Streptococcus</i> spp. Viridans Group	130	129	130	129	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with LHB	220	219	219	218	99.5%	99.7%
Total	528	524	524	522	99.6%	99.6%

Table 7. Summary Data and % Essential Agreement of *Streptococcus* spp. Clinical and Challenge Isolates Using the Auto Read Method

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Streptococcus pneumoniae</i>	151	151	148	148	98.0%	98.0%
<i>Streptococcus</i> spp. Viridans Group	104	104	104	104	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with LHB	195	194	194	193	99.5%	99.5%
Total	450	448	446	444	99.1%	99.1%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Streptococcus pneumoniae</i>	25	25	25	25	100.0%	100.0%
<i>Streptococcus</i> spp. Viridans Group	25	25	25	25	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with LHB	25	25	25	25	100.0%	100.0%
Total	75	75	75	75	100.0%	100.0%

Organism Group	Number of Isolates		Essential Agreement		% Essential Agreement	
	All	Evaluate	Total	Evaluate	Total	Evaluate
<i>Streptococcus pneumoniae</i>	176	176	173	173	99.0%	99.0%
<i>Streptococcus</i> spp. Viridans Group	129	129	129	129	100.0%	100.0%
<i>Streptococcus</i> spp. Beta-hemolytic Group with LHB	220	218	219	217	99.8%	99.8%
Total	525	523	521	519	99.2%	99.2%

Results

Essential agreement for telavancin on the Sensititre® susceptibility plate compared to the CLSI reference microdilution plate was calculated for each method (automated and manual read) using the +/- one log₂ dilution standard. The calculation for Evaluable excluded any test results where MICs were off-scale for the dilutions tested. Essential agreement rates are shown for Gram positive isolates in tables 4 and 5, and for *Streptococcus* spp. isolates in tables 6 and 7.

Clinical Isolates and CDC Challenge Organisms

• **Gram positive Isolates:** The overall essential agreement for telavancin (+/- one log₂ dilution) was 99.6% for the manual method and 99.0% for the automated method (Tables 4 and 5).

• **Streptococcus spp. Isolates:** The overall essential agreement for telavancin (+/- one log₂ dilution) was 99.6% for the manual method and 99.2% for the automated method (Tables 6 and 7).

Interlaboratory Reproducibility

• **Gram positive Isolates:** Reproducibility testing results for telavancin (+/- one log₂ dilution) from the modal MIC was 100% for the automated method and 100% for the manual read method (Table 8).

• **Streptococcus spp. Isolates:** Reproducibility testing results for telavancin (+/- one log₂ dilution) from the modal MIC was 99% for the automated method and 97% for the manual read method (Table 8).

Table 8. Interlaboratory Reproducibility % Essential Agreements +/- one log₂ Dilution of the Modal MIC for Telavancin

	Auto gram-positive	Manual gram-positive	Auto streptococcus spp.	Manual streptococcus spp.
Between-site total isolates tested	75	75	75	75
Between-site isolates within +/- 1 well from mode	75	75	74	73
Between-site reproducibility ratio	75/75	75/75	74/75	73/75
Between-site reproducibility %	100%	100%	98%	97%
Total essential agreement	72	73	73	73
Essential agreement %	96%	97%	97%	97%

Conclusions

This study validates that the Sensititre® 18-24 hour susceptibility system (both automated and manual read) demonstrated an equivalent level of performance compared to the CLSI M07 reference broth microdilution plate when testing telavancin against Gram-positive and *Streptococcus* spp. clinical and challenge isolates. The high level of essential agreement obtained by the Sensititre® 18-24 hour susceptibility method and the CLSI reference method suggests that this is an acceptable method for susceptibility testing of telavancin.

References

Clinical and Laboratory Standards Institute. 2009. *Methods for dilution antimicrobial susceptibility tests for bacteria that grow aerobically; approved standard-eighth edition. Approved document M07-A8.* Wayne, PA: CLSI.

Clinical and Laboratory Standards Institute. 2009. *Performance standards for antimicrobial susceptibility testing, 19th informational supplement M100-S19.* Wayne, PA: CLSI.

Acknowledgement

The research and publication process was supported jointly by Theravance, Inc. and Astellas Pharma Global Development, Inc.